

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFD

Date: 10/18/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 27270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:49
INTERNAL STANDARD	VERIFIED	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:52
BLANK TEST	0.00	23:53

