

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/08/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05

