

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 01/04/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:41
INTERNAL STANDARD	VERIFIED	05:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:44
BLANK TEST	0.00	05:45

