

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFG

Date: 01/11/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:31
INTERNAL STANDARD	VERIFIED	00:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:34
BLANK TEST	0.00	00:35

