

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 01/25/2023

OPERATOR NAME:  
NICHOLAS A MARCANO  
Permit #: 49335  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:25
INTERNAL STANDARD	VERIFIED	06:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:26
BLANK TEST	0.00	06:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:27
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:28
BLANK TEST	0.00	06:29

CFO

