

SUPERVISOR TEST

CFD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 11/22/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:44
INTERNAL STANDARD	VERIFIED	01:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48

