

SUPERVISOR TEST

LFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/29/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:33
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:35
BLANK TEST	0.00	07:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37

