

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 12/13/2023

OPERATOR NAME:
ROBERT J PRODY
Permit #: 50912
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46

