

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: OC50

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
7/19/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
7/27/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIRA	NO
8/2/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	YES
8/9/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO
8/16/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIRA	NO
8/24/23	23190	34.01	34.01	34.01	MP4189	.09	.10	.10	DIXON	NO
8/30/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.09	DIXON	NO
9/1/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIRA	YES
9/6/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
9/14/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO
9/20/23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
9-27-23	23190	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	NO

\* - Please enter a ✓ to denote simulator solution change taking place

CFO