

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: DCSO

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/16/23	23030	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	No
5/17/23	23030	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	No
5/23/23	23030	34.01	34.01	34.01	MP4189	.10	.10	.10	OSICK	No
6/1/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	YES
6/7/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	No
6/14/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	No
6/21/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	OSICK	NO
6/28/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	MARCANO	NO
7/5/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO
7-2-23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	Yes
7/5/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	OSICK	NO
7/12/23	23090	34.01	34.01	34.01	MP4189	.10	.10	.10	DIXON	NO

\* - Please enter a √ to denote simulator solution change taking place

CFO