

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 04/01/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:02
INTERNAL STANDARD	VERIFIED	01:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:03
BLANK TEST	0.00	01:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:04
BLANK TEST	0.00	01:05

