

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/07/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:39
INTERNAL STANDARD	VERIFIED	04:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:39
BLANK TEST	0.00	04:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:40
BLANK TEST	0.00	04:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:41
BLANK TEST	0.00	04:42

