

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 04/17/2024

OPERATOR NAME:  
NICHOLAS A MARCANO  
Permit #: 49335  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:10
INTERNAL STANDARD	VERIFIED	06:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:12
BLANK TEST	0.00	06:13

CFO

