

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/24/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:53
INTERNAL STANDARD	VERIFIED	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56

