

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 05/22/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:30
BLANK TEST	0.00	02:31

