

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/26/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:32
INTERNAL STANDARD	VERIFIED	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:33
BLANK TEST	0.00	17:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:34
BLANK TEST	0.00	17:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:35
BLANK TEST	0.00	17:36

