

SUPERVISOR TEST

CFU

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 07/03/2024

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 05:21 |
| INTERNAL STANDARD       | VERIFIED | 05:22 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:22 |
| BLANK TEST              | 0.00     | 05:23 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:23 |
| BLANK TEST              | 0.00     | 05:24 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:24 |
| BLANK TEST              | 0.00     | 05:25 |

