

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/24/2024

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:49
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:50
BLANK TEST	0.00	19:51

