

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/15/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:59
INTERNAL STANDARD	VERIFIED	19:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:59
BLANK TEST	0.00	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:00
BLANK TEST	0.00	20:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:02
BLANK TEST	0.00	20:02

