

Date: 10/23/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

CFO

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:48
INTERNAL STANDARD	VERIFIED	20:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:51
BLANK TEST	0.00	20:52

