

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 10/30/2024

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:47
INTERNAL STANDARD	VERIFIED	17:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:48
BLANK TEST	0.00	17:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:49
BLANK TEST	0.00	17:50

