

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 11/06/2024

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:36
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40

