

SUPERVISOR TEST

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 11/14/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:47
INTERNAL STANDARD	VERIFIED	21:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:50
BLANK TEST	0.00	21:51

