

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/20/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21

