

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/28/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:22
INTERNAL STANDARD	VERIFIED	19:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25

