

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFU

Date: 12/04/2024

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:43
INTERNAL STANDARD	VERIFIED	03:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:45
BLANK TEST	0.00	03:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:47

