

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 12/11/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:19
BLANK TEST	0.00	08:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:22
BLANK TEST	0.00	08:23

