

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 12/18/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:04
BLANK TEST	0.00	00:05

