

SUPERVISOR TEST

CRD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 12/25/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:56
INTERNAL STANDARD	VERIFIED	22:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:57
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:59
BLANK TEST	0.00	23:00

