

**Infrared Instrument**

**Weekly Calibration Verification Test Record**

Department: OC 50

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
7/10/24	24140	34.01	34.01	34.01	M74189	.09	.09	.09	OSIKA	N
7/11/24	24140	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	N
7/25/24	24140	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	N
7/30/24	24140	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	N
8/3/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	Y
8/12/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	N
8/15/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	Eilers	N
8/21/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	OSIKA	N
8/28/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	Eilers	N
9/1/24	24190	34.01	34.01	34.01	M74189	.09	.09	.09	DIXON	Y
10/1/24	24270	34.01	34.01	34.01	M74189	.10	.09	.10	Eilers	Y
10-9-24	24270	34.01	34.01	34.01	M74189	.09	.09	.09	Eilers	N

\* - Please enter a ✓ to denote simulator solution change taking place

CF0