

**Infrared Instrument**

**Weekly Calibration Verification Test Record**

Department: OC50

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/23/24	24270	34.01	34.61	34.61	MP4189	.09	.09	.09	E. I. S	N
10/30	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	Dixon	N
11/1/24	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	DIXON	Y
11/6/24	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	Dixon	N
11/14/24	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
11/20/24	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
11/28/24	24270	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
12/2/24	24350	34.01	34.01	34.01	MP4189	.09	.09	.10	Carraon	Y
12/4/24	24350	34.01	34.01	34.01	MP4189	.09	.09	.09	Dixon	N
12/11/24	24350	34.01	34.01	34.01	MP4189	.09	.09	.09	Cusack	N
12/18/24	24350	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
12/25/24	24350	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N

\* - Please enter a ✓ to denote simulator solution change taking place

CFO