

Infrared Instrument

Weekly Calibration Verification Test Record

Department: OCSD

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/1/24	24040	31.01	34.00	31.01	MP4189	.10	.10	.10	DIXON	Y
5/7/24	24040	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIKA	N
5/15/24	24040	31.01	31.01	37.01	MP4189	.10	.10	.10	DIXON	N
5/22/24	24040	34.01	34.01	34.01	MP4189	.09	.09	.10	OSIKA	N
5/30/24	24040	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
6-2-24	24140 24140	34.01	34.01	34.01	MP4189	.09	.09	.09	Eilers	Y
6/5/24	24140	31.01	31.01	31.01	MP4189	.09	.09	.09	DIXON	N
6/12/24	24140	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
6/19/24	24140	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	N
6/26/24	24140	31.01	31.01	31.01	MP4189	.09	.09	.09	DIXON	N
7/1/24	24140	34.01	34.01	34.01	MP4189	.09	.09	.09	OSIKA	Y
7/4/24	24140	31.01	31.01	31.01	MP4189	.09	.09	.09	DIXON	N

* - Please enter a √ to denote simulator solution change taking place

CFO