

Infrared Instrument

Weekly Calibration Verification Test Record

Department: OC50

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
12/6/23	23350	3401	3401	3401	MP4189	.10	.10	.10	OSIKA	N
12/13/23	23350	3401	3401	3401	MP4189	.10	.09	.10	Prody	N
12/20/23	23350	3401	3401	3401	MP4189	.09	.09	.10	OSIKA	N
12/27/23	23350	3401	3401	3401	MP4189	.09	.09	.09	DIXON	N
1/1/24	23350	3401	3400	3401	MP4189	.10	.10	.10	OSIKA	Y
1/3/24	23350	3401	3401	3401	MP4189	.10	.10	.10	OSIKA	N
1/10/24	23350	3401	3401	3401	MP4189	.10	.09	.10	DIXON	N
1/17/24	23350	3401	3401	3401	MP4189	.09	.09	.09	OSIKA	N
1/23/24	23350	3401	3401	3401	MP4189	.09	.09	.09	DIXON	N
1/31/24	23430	3401	3401	3401	MP4189	.10	.10	.10	OSIKA	Y
2/7/24	23430	3401	3401	3401	MP4189	.09	.09	.10	DIXON	N
2/14/24	23430	3401	3401	3401	MP4189	.09	.09	.10	DIXON	N

\* - Please enter a ✓ to denote simulator solution change taking place

CFO