

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:45
INTERNAL STANDARD	VERIFIED	05:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:45
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:48
BLANK TEST	0.00	05:49

