

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 12/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:06
INTERNAL STANDARD	VERIFIED	14:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:08
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:09
BLANK TEST	0.00	14:10

