

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/19/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:45
INTERNAL STANDARD	VERIFIED	04:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:46
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:47
BLANK TEST	0.00	04:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:48
BLANK TEST	0.00	04:49

