

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/01/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

CFO

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:32
INTERNAL STANDARD	VERIFIED	00:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35

