

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:27
INTERNAL STANDARD	VERIFIED	15:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:29
BLANK TEST	0.00	15:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:30
BLANK TEST	0.00	15:31

