

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/25/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:21
INTERNAL STANDARD	VERIFIED	17:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:22
BLANK TEST	0.00	17:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:23
BLANK TEST	0.00	17:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:24
BLANK TEST	0.00	17:25

