

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/25/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25

