

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/05/2025

OPERATOR NAME:
ANTHONY J TRAVIGLIA
Permit #: 53024
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:44
INTERNAL STANDARD	VERIFIED	19:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:44
BLANK TEST	0.00	19:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:45
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:47
BLANK TEST	0.00	19:48

