

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/26/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:27
INTERNAL STANDARD	VERIFIED	11:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:27
BLANK TEST	0.00	11:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:28
BLANK TEST	0.00	11:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:30
BLANK TEST	0.00	11:31

