

SUPERVISOR TEST

CF8

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/09/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:57
INTERNAL STANDARD	VERIFIED	05:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:58
BLANK TEST	0.00	05:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:59
BLANK TEST	0.00	06:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:00
BLANK TEST	0.00	06:01

