

# SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 10/22/2025

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 05:17 |
| INTERNAL STANDARD       | VERIFIED | 05:17 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:18 |
| BLANK TEST              | 0.00     | 05:19 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:19 |
| BLANK TEST              | 0.00     | 05:20 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 05:20 |
| BLANK TEST              | 0.00     | 05:21 |

