

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 01/08/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:12
INTERNAL STANDARD	VERIFIED	18:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:14
BLANK TEST	0.00	18:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:15
BLANK TEST	0.00	18:16

