

SUPERVISOR TEST

CFD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 01/22/2025

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:49
INTERNAL STANDARD	VERIFIED	17:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:49
BLANK TEST	0.00	17:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:51
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:52
BLANK TEST	0.00	17:53

