

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 01/29/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:09
INTERNAL STANDARD	VERIFIED	00:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13

