

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/26/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:12
INTERNAL STANDARD	VERIFIED	02:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16

