

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CEP

Date: 03/05/2025

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:23
INTERNAL STANDARD	VERIFIED	20:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:23
BLANK TEST	0.00	20:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:24
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:25
BLANK TEST	0.00	20:26

