

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/23/2025

OPERATOR NAME:
DAN J CORCORAN
Permit #: 53797
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:06
INTERNAL STANDARD	VERIFIED	23:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:07
BLANK TEST	0.00	23:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:10

